Fold Fold Fold

91 mm 101 mm 101 mm 91 mm

178 mm

Outside

Garantía Limitada de por Vida

Calidad es el sello de todos de los productos de Lumex®. De confianza por profesionales de cuido de salud y usuarios por más de 50 anos, Lumex tiene una linea completa de asistencia para pacientes y asientos especializada de asistencia sanitaria.

Los productos de Lumex son diseñados y hechos a estándares precisos y son revisados cuidadosamente antes de envió. La Garantía Limitada de por Vida es un expresión de nuestro confianza en los materiales y factura de nuestros productos. Somos dedicados a proveer productos que pueden dar anos de servicio fiable a nuestros clientes.

GF Health Products, Inc. garantía las partes específicos de estos modelos de Lumex escritos en el forro de este “hangtag” para la vida del producto sobre ser libre de defectos en materiales y factura cuando usado por el propósito entendido, sobre condiciones normales, y próvido que el producto reciba cuido apropiado.

Partes cubiertos debajo de esta garantía incluye: marco, soldadura, tornillos, componentes de posicionar el marco. Artículos específicos “excluidos” de esta garantía de cual quiere tipo: “TODAS” otras partes no mencionados especificamente arriba. La garantía no se extenderá a componentes no durables como accesorios punta, cuales son sujetos a desgasto normal y requieren remplace periódico. Esta garantía no incluye cargos de labor incurridos con instalación de partes de reemplazo o cargos de transporte o carga asociados al fabricante.

Si hay un defecto en el producto que se cubra por la garantía, nosotros, a nuestra opción, podemos reparar o reemplazar el producto.

Las garantías que figuran en este documento contiene todas las declaraciones y garantías en relación con la materia objeto de este documento, y sustituyen todas las negociaciones previas, acuerdos y entendimientos con respecto a la misma. El receptor del presente documento reconoce y declara que no ha basado en ninguna declaración, afirmación, garantía, contrato de garantía u otra promesa, excepto los establecidos en este documento.

Para información adicional de este producto de Lumex o esta garantía, por favor de llamar a su distribuidor autorizado de GF Health Products, Inc.

Back Cover

LUMEX®
A Graham-Field Brand

4mm Hole Drill

- Improves mobility and allows for greater independence for the user on the go
- Lightweight design makes rollator easy to use, fold and store for convenient transport
- Padded seat and backrest provide additional comfort
- Ergonomic hand grips and hand brakes allow user to operate with ease
- Complete with basket for storage of personal items
- 6" wheels are ideal for indoor and outdoor use
- RJ4302 model has a lower seat position for shorter individuals
- 300 lb weight capacity (350 lb on RJ4318 Wide models)
- Available in various attractive colors
- Newly added Reflectors increase visibility

- Mejora movilidad y provee mas independencia al usuario siempre en la calle
- Diseño ligero hace que el rodador es fácil de usar, doblar y almacenar para transporte conveniente
- Asiento y respaldo acolchado provee mas comodidad
- Agarres de mano ergonómicas y frenos de mano permite a usuario operarlo con facilidad
- Viene completo con canasta para almacenaje de cosas personales
- Ruedas de 6" son ideal para uso adentro y afuera
- RJ4302 tiene una posición de asiento bajo para individuales mas pequeños
- Capacidad de peso de 300 lb (350 lb en models anchos de RJ4318)
- Disponible en varias colores atractivos
- Reflectores nuevos aumentan visibilidad

For the most updated and current information on packaging, warranties, products and specifications,please visit our website at www.grahamfield.com

Para la información más ultimo sobre embalaje, garantías, productos y especificaciones, por favor de visitar a nuestro sitio de internet a www.grahamfield.com

© 2012, GF Health Products, Inc.
2935 Northeast Parkway, Atlanta, Georgia 30360.
Made in China. www.grahamfield.com

GF06000057RevD12

Front Cover

6mm 6mm

LUMEX®
A Graham-Field Brand

New Wider Version!

Walkabout Lite, Hemi and Wide
4-Wheel Rollator

Walkabout Lite, Hemi and Ancho Rodador de 4 Ruedas
Items RJ4300, RJ4302 & RJ4318

User Instructions and Warranty Information
Important: Read Before Use
Instrucciones e Informacion de la Garantía
Importante: Lea Antes de Uso

GRAHAM FIELD

Lifetime Limited Warranty

Quality is the hallmark of all Lumex® products. Trusted by healthcare professionals and end-users for over 50 years, Lumex features a full line of patient aids and specialty healthcare seating.

Lumex products are designed and built to precise standards and are carefully inspected prior to shipment. This Lifetime Limited Warranty is an expression of our confidence in the materials and workmanship of our products. We are dedicated to providing products that will give our customers years of dependable service.

GF Health Products, Inc. hereby warrants specific parts of the Lumex models listed on the cover of this hangtag for the lifetime of the product against defective workmanship and/or material defects when used for the purpose intended, under normal conditions, and provided the product receives proper care.

Parts covered under the lifetime warranty include: frame, welding, screws, and frame locking components. Specific items “excluded” from this warranty of any kind are: “ALL” other parts not specifically listed above. The warranty does not extend to non-durable components, such as tips, which are subject to normal wear and require periodic replacement. The warranty does not include any labor charges incurred in replacement part(s) installation or any associated freight or shipping charges to the manufacturer.

Should a product defect that is covered by this warranty arise, we will, at our option, repair or replace the product.

The warranties contained herein contain all the representations and warranties with respect to the subject matter of this document, and supersede all prior negotiations, agreements and understandings with respect thereto. The recipient of this document hereby acknowledges and represents that it has not relied on any representation, assertion, guarantee, warranty, collateral contract or other assurance, except those set out in this document.

For additional information on this Lumex product or this warranty, please contact an authorized GF Health Products, Inc. Distributor.

Also available from the Lumex line of Patient Aids.....

También disponible de la línea de Asistencias para Pacientes de Lumex

Low Profile Quad Canes
Bastones con Base Cuadrada de Perfil Bajo
Item 6118A shown here

Three-in-One Commodes
Inodoros de Tres-en-Uno
Item 7103A-4 shown here

Maxi-Drain Transfer Benches
Bancos de Transferir de Drenaje Maximo
Item 782A shown here

Available accessories for Walkers
For other options please visit us online at www.gfahamfield.com

Walker Pouch
Lumex Bolsa del Andador
Item 6032D4

Walker Tray
Bandeja Para Andador
Item 6039004

Walker Basket
Canasta de Andador
Item 6039044

Glide Skis
Esqui Deslizante
Item 6032108

Inside

Instructions for Use – Please Read All Instructions Before Use.

Please consult a healthcare professional for the proper adjustment.

Opening the Rollator

1. Remove Rollator and other components from carton.
2. Unfold frame; spread front and back legs apart.
3. Push down locking mechanism to secure legs in place.
   NOTE: Ensure the side crossbars (locking mechanism) lock down.

Folding the Rollator

1. If necessary, remove the accessories and lift the seat.
2. Pull up on the Folding Pull Strap that is attached to crossbar (locking mechanism) underneath the seat.
3. Push the two handgrips together.

Using the Hand Brake

Brakes should be used in one of the following methods:

- To Remain Stationary: Hold the hand brake up.
- To Continue Mobility: Release the brake. To unlock, pull up on the top portion of the brake handle.
- To Sit on Seat: Push down on the bottom portion of the brake handle to lock.
- To unlock, pull up on the top portion of the brake handle.
  CAUTION: Always lock hand brakes when user is immobile; never sit on rest seat if hand brakes are not locked.

Adjusting the Height

1. Remove the adjustment knob and bolt by turning counterclockwise.
2. Consult with healthcare professional for proper adjustment.
3. Install the adjustment knob into one of the adjustment holes by turning clockwise.
4. Repeat STEPS 1– 3 for the other side.

Make sure that the height adjustment is at the same height for both handles to ensure safety and proper use.

NOTE: After adjusting the height, the brakes of the rollator must be checked and, if necessary, adjusted.

Attaching Basket, Pouch or Tote Bag

Basket
Install the basket underneath the seat. On some rollator models, the basket is designed to fit underneath the seat or in the front by positioning the hooks on the front seat support frame.

Pouch and Tote Bag
Hang the tote bag between the rollator frame tubes and secure with snap-button flaps.

Maintenance

- For dust and dirt, wipe Rollator gently with damp, soft cloth.
- Inspect wheels periodically for wear or damage. Use a mild soap and water solution to clean, rinse with clear water, then dry with a soft cloth.
- Never use organic solvents, abrasive cleaners, or scouring pads on any part of the rollator.
- Verify operation of the brakes and have them adjusted if necessary by your GF Health Products, Inc. authorized distributor.
- Inspect rollator weekly or on a set routine preventative maintenance schedule for normal wear and tear. If any component is not in proper working order or appears worn, contact your GF Health Products, Inc. distributor for repair and/or appropriate replacement parts.
- If handgrip or hand brakes are loose, **DO NOT** use. Contact your GF Health Products, Inc. authorized distributor.
- Replace any broken, damaged or worn items immediately.
- For the best service from your rollator, always use replacement parts from GF Health Products, Inc.
- Periodically inspect the casters and caster stems for assembly tightness and verify that the wheels are free of hair or other debris that could interfere with free wheel operation.

Item Number	*Weight Capacity*
RJ4300s	*300 lb*
RJ4318s	*350 lb*
RJ4302s	*300 lb*
RJ4805s	*300 lb*
RJ4400s	*500 lb*
RJ4405s	*500 lb*
RJ4402s	*500 lb*
609201s	*300 lb*

Safety Guidelines – Please Read Before Use

1. Each individual should ALWAYS consult with their physician or other health care professional to determine proper adjustment and usage.
2. The brakes MUST be in the locked position before using the seat.
3. All wheels MUST be in contact with the floor at all times during use. This will ensure the rollator is properly balanced.
4. This product has a maximum user weight limitation of 300 lbs (350 lb on RJ4318 Wide models).
   ALWAYS observe the user weight limitation on the hangtag of this product and on the chart referenced at left.
5. Ensure handgrips DO NOT twist on rollator handle - otherwise damage or injury may occur.
6. ALWAYS test to see that the rollator and attachments are properly and securely locked in the open position BEFORE using.
7. After installation and before use, ensure that all attaching hardware is securely tightened.
8. ALWAYS use caution when using the rollator on wet and slippery surfaces. DO NOT use the rollator on stairs and steep inclines.
9. DO NOT hang anything from the frame of the rollator. Items should be placed in the basket or tote bag.
10. The rollator basket has a weight limitation of 11 lbs (5 kg). The tote bag has a weight limitation of 10 lbs (5 kg). Items placed in either the basket or the bag should not protrude from the basket or bag.
11. The backrest should ALWAYS be in place and is NOT designed to support the total body weight of the user.
12. Before attempting to reach objects or pick them up from the floor by reaching down between your knees, place feet securely on the floor. Use EXTREME caution when reaching for any object.
13. GF Health Products, Inc. assumes no responsibility for any damage or injury caused by improper installation, assembly or use of this product.
14. Inspect rollator weekly or on a set routine preventative maintenance schedule for normal wear and tear. If any component is not in proper working order or appears worn, contact your Lumex dealer for repair and/or appropriate replacement parts.
15. If the hand brakes fail to stop the Rollator, contact your GF Health Products, Inc. authorized distributor immediately for adjustment or replacement.

WARNING: Indicates a potential hazard situation or unsafe practice that, if not avoided, could result in death or serious personal injury.

CAUTION: Indicates a potential hazard or unsafe practice that, if not avoided, could result in personal injury or product/property damage.

NOTE: Provides application recommendations or other useful information to ensure that you receive the greatest benefit from this product.

WARNING

WARNING: DO NOT use this product without first completely reading and understanding these instructions and any additional instructional material supplied with this product. If you are unable to understand the warnings, cautions or instructions, contact a healthcare professional, dealer or technical personnel before attempting to use this equipment - otherwise, injury or damage may occur.

WARNING: DO NOT use the rollator as a wheelchair or transport device. Rollators are not intended to be propelled while seated.

WARNING: If components are damaged or missing, contact your distributor immediately. DO NOT substitute parts.

WARNING: Before use, ensure that all assembled parts are secure and all hardware is tightened.

WARNING: Care should be taken to ensure that all hand and height adjustments are secure, and that wheels and moving objects are in good working order before using this rollator.

WARNING: When using the rollator in a stationary position, the hand brakes MUST be locked.

WARNING: ALWAYS observe the user weight limitation on the hangtag of this product and on the chart referenced at left.

WARNING: DO NOT attempt to reach objects if you have to move forward in the seat. Reaching for these objects will cause a change to the weight distribution of the rollator and may cause the rollator to tip, resulting in injury or damage.

Inside Front Cover

Instructions for Use – Please Read All Instructions Before Use.

Por favor de consultar con un profesional de cuido de salud para ajustamiento apropiado.

Abriendo el Rodador

1. Remueva el Rodador y los otros componentes del cartón.
2. Desdobla el marco; extiende las patas delanteras y traseras.
3. Aprieta hacia abajo en el mecanismo de cerrar para asegurar las patas en posición.
   Nota: Asegura que los barras de los lados (mecanismo de cerrar) están posicionados.

Doblando el Rodador

1. Si necesario, remueva los accesorios y levanta el asiento.
2. Jale hacia arriba en el Cordón de Doblar que esta fijado en la barra debajo del asiento.
3. Apreté los mangos juntos.

Usando el Freno de Mano

Los frenos deben ser usados en uno de los siguientes métodos:

- Para Mantener estacionario: Aguante el freno hacia arriba.
- Para Continuar Movilidad: Suelta el freno.
- Para Sentarse: Apreté hacia abajo en la porción abajo del freno para cerrar.
- Para Abrir, jale en la porción superior del freno de mano.

  Caución: Siempre cierre los frenos de mano cuando el usuario es inmóvil; nunca se asiente en el asiento si los frenos de mano no están asegurados.

Para Ajustar la Altura

1. Remueve la perilla y tornillo de ajustamiento en sentido contrario a las agujas del reloj.
2. Consuite con un profesional de cuido de salud para ajustamiento apropiado.
3. Instale la perilla de ajustamiento en uno de los huecos en sentido a las agujas del reloj.
4. Repite PASOS 1– 3 para el otro lado.

Asegure que el ajustamiento de altura es lo mismo para los dos mangos para asegurar seguridad y uso apropiado.

Nota: Después de ajustar la altura, los frenos del rodador deben ser examinados y si necesario, ajustados.

Poniendo la Canasta, Bolsa y el Bolsón

Canasta
Instala la canasta debajo del asiento. En algunos de los modelos, la canasta es diseando para poner debajo del asiento o al frente posicionado con los ganchos en la parte del frente del marco.

Bolsa y Bolsón
A gancha el bolsón dentro del marco del rodador y asegura con los solapas.

Mantenimiento

- Para polvo, limpie el rodador con una toalla suave y mojada.
- Examine las ruedas por daño y desgasto periódicamente. Use un jabón suave y agua para limpiar, lave con agua limpia y seca con una toalla suave.
- Nunca use solventes orgánicos, limpiadores abrasivos o estropajos en cual quiere partedel rodador.
- Verifica la operación de los frenos y permite que su distribuidor de GF Health Products, Inc. se los ajusta si es necesario
- Examine su rodador seminal o en un programa fija de mantenimiento preventiva para función y desgasto. Si algunos componentes so están funcionando correctamente o se ven con desgasto, llame a su distribuidor de GF Health Products, Inc. para reparación y partes de reemplazo.
- Si los mangos o frenos de mano están sueltos, NO LOS use. Llame a su distribuidor autorizado de GF Health Products, Inc.
- Para el mejor servicio de su rodador, siempre use partes de reemplazo de GF Health Products, Inc.
- Periódicamente examine los rueditas y vástagos para su apriete y verifica que las ruedas están libres de pelo o otros escombros que pueden interferir con el operación libre de la rueda.

Numero de modelo	*Capacidad de Peso*
RJ4300s	*300 lb*
RJ4318s	*350 lb*
RJ4302s	*300 lb*
RJ4805s	*300 lb*
RJ4400s	*500 lb*
RJ4405s	*500 lb*
RJ4402s	*500 lb*
609201s	*300 lb*

Inside Back Cover

Directriz de Seguridad – Por Favor de Leer Antes de Uso

1. Cada individual SIEMPRE debe consultar con su doctor o otro profesional de cuido de salud para determinar ajustamiento y uso apropiado.
2. Los frenos TIENEN que estar en la posición asegurados antes de usar el asiento.
3. Todas las ruedas TIENEN que estar en contacto con el piso todo el tiempo cuando en uso. Esto puede asegurar que el rodador esta balanceado apropiadamente.
4. Este producto tiene una limitación de peso máximo de usuario de 300 lb. (350 lb en los modelos de RJ4318 Ancho)
   SIEMPRE tiene que observar la limitación de peso en el “hangtag” de este producto o en la tabla a la izquierda.
5. Asegure que los mangos NO SE dan vuelta en el rodador – o si no, daño o herida puede ocurrir.
6. SIEMPRE prueba que el rodador y accesorios están puestos seguramente en la posición abiertos ANTES de uso.
7. Después de instalación y antes de uso, asegura que el ferretero esta seguramente apretados.
8. SIEMPRE use caución cuando usando los rodadores en superficies mojados y resbalosos. NO USE el rodador en escaleras o inclinas.
9. NO fija cual quiere cosa del marco del rodador. Cosas deben ser puestos en la canasta o el bolsón.
10. La canasta del rodador tiene una limitación de peso de 11 lb (5 kg). El bolsón tiene una limitación de 10 lb (5 kg). Cosas puestas en la canasta o bolsón no DEBEN sobresalir.
11. El respaldo debe SIEMPRE estar en posición y no es designado soportar todo el peso del cuerpo del usuario.
12. Antes de atemptar a alcanzar objetos o recojerlos del piso alcanzando dentro sus rodillas, asegura sus pies en el piso. Usa caucion EXTREMO cuando alcanzando por cual quiere objeto.
13. GF Health Products, Inc. no asume ninguin responsabilidad por cual quiere daño o herida causado por instalación, asemblaje o uso incorrecto de este producto.
14. Examine su rodador seminal o en un programa fija de mantenimiento preventiva para función y desgasto. Si algunos componentes no están funcionando y se ve con desgasto, llame a su distribuidor de Lumex para reparación o reemplace de partes.
15. Si los frenos de mano fallan o y parran el rodador, llame a su distribuidor autorizado de GF Health Products, Inc. para ajustamiento o reemplazo.

ADVERTENCIA: Indica un riesgo potencial, situación o practica insegura que si no se evita puede resultar en muerte o herida personal severa.

CAUCION: Indica una amenaza potencial o practica insegura, que si no es evitada podría resultar en daño personal o a la propiedad.

NOTA: Provee recomendaciones u otra información para asegurar que usted reciba lo mejor de su producto.

ADVERTENCIA

ADVERTENCIA: NO DEBE usar este producto sin leer y entender completamente estas instrucciones y material adicional que viene con este producto. Si no puedes entender los advertencias, cauciones o instrucciones, llame a su profesional, distribuidor o personal técnicos antes de tratar de usar este equipo – o si no, herida o daño puede ocurrir.

ADVERTENCIA: NO DEBE usar el rodador como una silla de ruedas o silla de transporte. Rodadores son entendidos para propulsado mientras sentado.

ADVERTENCIA: Si los componentes están dañados o si faltan, llame a su distribuidor inmediatamente. NO substituye partes.

ADVERTENCIA: Antes de uso, asegure que las partes instalados están seguros y todo el ferreteria esta apretado.

ADVERTENCIA: Cuido debe estar hecho para asegurar que todos los ajustamientos de mano y altura están seguros, y que las ruedas y objetos que se mueven están en buenas condiciones antes de usar el rodador.

ADVERTENCIA: Cuando usando el rodador en la posición estacionado, los frenos de mano TIENEN que estar asegurados.

ADVERTENCIA: SIEMPRE observa la limitación de peso del usuario en el “hangtag” de este producto o en la tabla abajo.

ADVERTENCIA: NO intente alcanzar para objetos si tiene que moverse hacia adelante en el asiento. Alcanzando para estos objetos puede causar un cambio en la distribución de peso del rodador y puede causar que el rodador se dobla, resultando en herida o daño.

384 mm X 356 mm
Hantag to be printed with 4 Color High Gloss Ink plus UV coating on 80 LB. White Stock.
Red guides are to indicate folds and layout only. They do not print.
Score all folds
Drill holes where indicated, 4mm
Use a CLEAR 25.4 MM round perferated waffer to seal hangtag